Light Pediatric Bath Chair

Model : BCL001 Series, BCL002 Series, BCL003 Series, BCL004 Series,
LTS Series, LSS Series



BCL001



BCL003



BCL004

Version: A/0 Release date: 19.01.2026

Important

These instructions provide important information for the safe use and maintenance of the Light Bath Chair. Please take the time to read this informative instruction before using this product. It is very important to follow the instructions, if not used in accordance with the instructions may cause serious injury or death to the user.

Intended Use

The Light is a bath chair designed to support children and young adults during bathing or showering in a home bathroom environment only.
It is not intended for outdoor use or long-term exposure to outdoor conditions. Prolonged exposure to moisture, salt, chemicals, or other corrosive substances may cause corrosion, material aging, or structural damage, which can compromise product safety.

Package Contents

The bath chair is completely assembled, you will only need to position the backrest, calf support and bath chair legs appropriately before use. Please check that you have all the accessories you ordered.

CE MD NON STERILE

Item Dimensions

Item Code	BCL001/BCL002	BCL001/BCL002 /BCL003	BCL001/BCL002 /BCL003	BCL001/BCL002 /BCL003	BCL001/BCL002 /BCL003
Item Dimensions	**10-XXXS**	**12-XXS**	**14-XS**	**16-S**	**16-M**
Back Height	38 cm	49 cm	60 cm	49 cm	60 cm
Seat Depth	20 cm	28 cm	35 cm	28 cm	35 cm
Seat Width	25.4 cm	30.5 cm	35.5 cm	43 cm	43 cm
Calf Length	20 cm	25 cm	29.5 cm	25 cm	29.5 cm
Overall Length Flat	78 cm	103 cm	126 cm	103 cm	126 cm
Overall Width	38 cm	44 cm	37.5 cm	45 cm	45 cm
Seat Height above Floor	15-40 cm	15-40 cm	15-46 cm	15-40 cm	15-46 cm
Weight Capacity	154.3 lbs.(70kg)	154.3 lbs.(70kg)	154.3 lbs.(70kg)	160 lbs.(72kg)	160 lbs.(72kg)

Item Code	BCL001/BCL002 /BCL003	BCL003	BCL003	BCL003	BCL003
Item Dimensions	**16-L**	**18-XL**	**20-XXL**	**22-XXXL**	**24-XXXXL**
Back Height	71.5 cm	71.5 cm	71.5 cm	71.5 cm	71.5 cm
Seat Depth	40 cm	40 cm	40 cm	40 cm	40 cm
Seat Width	43 cm	47 cm	51 cm	55.8 cm	61 cm
Calf Length	34 cm	34 cm	34 cm	45.5 cm	45.5 cm
Overall Length Flat	147 cm	147 cm	147 cm	157 cm	157 cm
Overall Width	45 cm	51.5cm	54cm	59 cm	64 cm
Seat Height above Floor	16-51 cm	16-48.5cm	16-48.5cm	16-48.5 cm	16-48.5 cm
Weight Capacity	160 lbs.(72kg)	286.6 lbs.(130kg)	286.6 lbs.(130kg)	286.6 lbs.(130kg)	286.6 lbs.(130kg)

Item Code	BCL004		
Item Dimensions	**16-S**	**16-M**	**16-L**
Back Height	49 cm	60 cm	71.5 cm
Seat Depth	42 cm	42 cm	42cm
Seat Width	43 cm	43 cm	43 cm
Overall Length Flat	100 cm	123 cm	143 cm
Overall Width	45 cm	45 cm	45 cm
Seat Height above Floor	15-40 cm	15-46 cm	16-51 cm
Weight Capacity	160 lbs.(72kg)	160 lbs.(72kg)	160 lbs.(72kg)



Note: Due to different measuring angles and fabric elasticity, the measurements may vary by ±1 cm.

Warranty Statement

The manufacturer's warranty period is 3 years, from the date of purchase as indicated on the sales receipt. This warranty does not cover the product failure due to owner misuse or negligence. It also does not cover the consumable items such as rubber parts, cover and other optional accessories, which are subject to normal wear and tear and need periodic replacement.

⚠ Warning

• This product shall be assessed the appropriateness and safety of all equipment for each user by qualified professionals only.

• Babies and young children are at risk of drowning when being bathed. Use extra caution when operating the bath chair in a bathtub.

• Never leave the user unattended, even for a few moments. The carer shall at all times be in contact with the user.

• Never let the user stand up in the bath chair.

• Do not park or leave the bath chair on a rough and uneven ground, or on an incline.

• To prevent the risk of a bath chair catching fire, do not place it near an exposed heat source such as cigarettes, electricity, an open fire.

• To avoid injury, ensure that the user is kept away when unfolding and folding this product.

• Chest Belt can cause strangulation! All the accessories must be carefully adjusted for the comfort and security of the user.

• To avoid serious injury from falling or sliding on the bath chair,the user must always use the accessories at all times.

• Overloading, incorrect folding or the use of non-approved accessories may damage or break this bath chair. Maximum weight limits: all three sizes of bath chair is 160 lbs.

• The bath chair should be inspected regularly for signs of damage and wear. Do not use the bath chair if any part is broken, torn or missing. The damaged or worn parts must be repaired immediately.

Maintenance

This product should be inspected, maintained, cleaned and/or washed routinely. Discontinue the use of this bath chair if it becomes damaged, broken or disassembled. Contact your supplier for qualified repair or replacement. Do not alter or repair by unqualified people. Customers made or arranged for modifications or repairs to be made without first having received written permission from the manufacturer, the manufacturer's warranty will be void.

Disposal

When a product reaches the end of its lifetime, it should be separated by material type so that the parts can be recycled and disposed of safely.

Adjustments



BCL001/BCL003




BCL004

Warning

- Do not let the user play with this product.
- To avoid finger and limb entrapment, be careful when adjusting the bath chair.
- Ensure that all the locking devices are adjusted correctly and tightened before use.
- Adjust the bath chair to a comfortable level before placing the user in it. If the user feels uncomfortable, remove from the product as quickly as is safely possible.
- With the exception of the calf support recline adjustment, the bath chair should not be adjusted while it is in use or occupied.

Adjust the backrest（BCL001)





Rotate the backrest to the desired position, release both buttons until they click into place

Adjust the backrest（BCL003/BCL004)



Lift the D-ring on the back of the backrest to adjust the angle.



Stop pulling the D-ring once the backrest is at the desired angle, and ensure the bar is inserted into the hole in the connection part between the backrest and seat.

Adjustments

Adjust the seat（BCL004)




Lift the D-ring on the back of the seat to adjust the angle.



Stop pulling the D-ring once the seat is at the desired angle, and ensure the bar is inserted into the hole in the connection part between the backrest and seat.


Adjust the calf support（BCL001/BCL003)





Rotate the calf support to the desired position, release both buttons until they click into place

Adjust bath chair back leg






Rotate the back leg to the desired position, release both buttons until they click into place

Adjustments

Adjust bath chair back leg and front leg






Warning

The likelihood of a bath chair tipping over when the back and front legs are at or below 90 degrees.







Folding the bath chair









① ② ③

Installation of connector cover (if applicable)



①



②



③

1. Place the connector cover over the connectors on both sides.

2. Thread the elastic strap through the hole in the center of the connector.

3. Hook the elastic strap onto the bottom of the backrest tube and make sure it is securely fastened and will not slip off.

Fixing the Center Strap of the Bath Chair Cover



1. Thread the strap through the hole in the connector from the inner side.

2. Wrap the strap around the connector and pass it between the connector and the base tube.

3. Fasten the buckles on both sides of the strap underneath the bath chair and make sure the cover is securely installed.

Removing the bath chair cover



①



②



③

Note: If your bath chair cover is equipped with connector covers, first remove the elastic straps of the connector covers from the backrest tubes when removing the bath chair cover. Then unfasten all the fixing strap buckles and remove the bath chair cover.

Tub / Shower Stand




LTS Series



LSS Series

Assembling the tub/shower stand








To assemble, press and hold the button on the slender tube, then insert it into the larger tube until the button pops out of the hole in the thick tube. Once you've ensured the stand is securely locked and stable, it is ready for use.


Warning

If the legs of the tub/shower stand are at or below 90 degrees, there is a high risk of it tipping over.

Maximum weight limits: 200 lbs.



Folding the tub / shower stand




①



②

Attaching the bath chair to the tub/shower stand



Warning

The tub/shower stand is only to be used in combination with Light bath chair. The tub/shower stand is only intended for bathing, no other uses are permitted.

Check that the bath chair is properly attached and stable on the tub/ shower stand before usage. NEVER lift the tub/shower stand with a user in it.

Inspect the tub/shower stand for damage and wear regularly, and discontinue use if any part is broken, torn, or missing. Repair or replace any damaged or worn parts as soon as possible.

Shower stand castors

Warning

To prevent falls or injuries, always lock all the castors when placing or removing the user.



Cleaning

The tub / shower stand can be wiped clean with a moist cloth and a mild cleaning agent.

Accessories



Warning


- All accessories are not intended for prolonged periods of support.
- Ensure that all accessories are adjusted for comfort and security.
- Do not place the chest belt and/or the Pelvic Belt around a user's neck.

Attaching the accessories to the bath chair

Head Lateral Support



Chest Belt



Fixed Pelvic Belt（M/L size）



①



②

Leg Straps



①



②



③



④

Calf Straps



①



②



③



④

Cleaning

The bath chair frame can be wiped clean with a moist cloth and a mild cleaning agent.

The bath chair cover is detachable and machine washable:

Washing with warm water (40°C/100°F) and using a mild detergent in a washing machine.
Place the bath chair cover inside a laundry bag to prevent damage to it and washing machine/dryer.
Tumble dry with low heat or dry. Do not bleach or iron. Do not dry clean.

The bath chair accessories are washed by hand .

Washing with warm water (30°C / 85°F) and using a mild detergent, then leaving them to dry before putting it back on the bath chair.
Do not tumble dry. Do not dry clean. Do not bleach or iron.

Maintenance tips

After using the bath chair, it is important to rinse with water to remove any remaining residue from soap or other bathing products.
Leave the bath chair to dry unfolded, and fold it after it is fully dried up for storage.
If the air is humid, store your bath chair and its accessories in a comparatively dry place and leave them to dry unfolded to prevent mold.
Continued and extended exposure to sunlight can change the colour of materials.

Table of Symbols

CE mark	Caution	Do not bleach
Medical Device	Manufacturer	Do not iron
Read the Instructions before Use	For indoor use	Machine wash- water temperature 40°C
Single patient - multiple use	Authorized representative in the European Community	Non-Sterile

Manufacturer: DongGuan LeYuan Plastic & Metal Technology Co.,LTD
Address:2nd Baoyuan Road,Tangxia Town,Dongguan City, Guangdong Province,China
Tel:+86-076982131662

Company: SUNGO Europe B.V.
Address: Fascinatio Boulevard 522, Unit 1.7, 2909VA Capelle aan den IJssel, The Netherlands
Tel: +31(0)10 3034500；+31(0)2021 11106